



Industrie Service

CERTIFICATE

1 MANUFACTURER

Invacare Corporation
One Invacare Way
Elyria, Ohio
USA

2 CERTIFICATE NO.

P / 600005748 / 001

3 CERTIFICATE HOLDER

INVACARE AQUATEC GmbH
Alemannenstr. 10
88136 Isny
Germany

4 TESTING LABORATORY



TÜV SÜD Industrie Service GmbH
Westendstrasse 199
D-80686 München

5 LABORATORY REPORT

Nr.: VR.60000.5748.001
Datum: 28.09.2009

6 STANDARDS APPLICABLE

3/2007 MDS-Hi: 2007-03-14

7 DESCRIPTION OF EQUIPMENT

Product group: Transportable Oxygen
Concentrator for medical use

Type: TPO100B

Serial Number: 09GF041755

8 OPERATING CONDITION AND RESULT

Flow rate: **Sound pressure level:**

1,0 l/min 33 dB(A)
2,0 l/min 36 dB(A)

Pulse Settings: **Sound pressure level:**

1 33 dB(A)
2 32 dB(A)
3 36 dB(A)
4 37 dB(A)
5 39 dB(A)

9 COMMENT

no

10 CERTIFICATE VALID:

Place, Date: Munich, 2009-09-30

Dr. Ralf Schirmer

11 QUALITY ASSURANCE:

Place, Date: Munich, 2009-09-30

Tobias Fleckenstein